



5" BACK
5" FRONT

VEP: .25" from right.
 1.25" Suggested clear sealing area for H-100 and Excel
 7.25"


Unloading




1
Unscrew the white plunger rod.



2
Pull the white plunger rod back until it clicks.



3
Turn the blue handle to the left (Counterclockwise) 180° until it clicks.




4
DO NOT RECAP! Slide the cartridge towards the blue handle. Invert the holder to remove the cartridge. Discard the used cartridge in a sharps container.

Notes:


- Save the holder to use with the next medication cartridge as desired.
- For reuse of the holder between patients, wipe all surfaces with a disinfecting wipe (i.e., PDI Super Sani-Cloth® Germicidal Disposable Wipe or chemical equivalent). Do not reuse the holder if there is visible soiling.
- The CARPUJECT holder should be used ONLY with the CARPUJECT cartridge.

Product inquiries should be directed to Hospira, Inc., Lake Forest, IL 60045 USA

PAA118266 



List No. 2049-02 Rx only GTIN: 10304092049020
 CL12

CARPUJECT™ Syringe System




Optional Assisted Method for Opening SLIM-PAK™ Package

- Slide the square opening on the holder over the tip of the SLIM-PAK cap.
- Hold the SLIM-PAK stationary while turning the holder to break the tamper evident seal and open the SLIM-PAK.


Loading

- Grasp the holder with the open side facing you and pull the white plunger rod back until it clicks.
- Turn the blue handle to the left (Counterclockwise) 180° until it clicks.
- Insert the cartridge in the holder by placing the end with the plunger in the holder first and ensure the cartridge lays flat.
- Screw the white plunger rod onto the plunger until you feel a slight resistance.
- Turn the blue handle to the right (Clockwise) 180° until it clicks. Note: This step creates an open fluid pathway; therefore the drug must be used or discarded after this step.
- Remove and discard green luer cap. Do not touch the luer end of the syringe. Expel air as needed. Administer prescribed dose.

Client:	Hospira Inc.	Art Number:	#120526,0001	Chart Number:	Chart 2	Proof Date:	3/16/2016	PLEASE PROOF CAREFULLY—This proof and the color samples are your responsibility; check for accurate spelling, punctuation, any legal requirements and proper positioning of copy and art. Check the attached color samples and make sure they meet your approval. This Proof Indicates Approximate Color Only. For Actual Colors, Use Ink Swatches Provided on Proof Approval Sheet. NOTE: Keyline and features DO NOT PRINT. They are for position only. THIS IS YOUR FINAL OPPORTUNITY TO CHECK FOR CORRECT ART PREPARATION. WE CANNOT BE HELD RESPONSIBLE FOR ANY ERRORS THAT HAVE NOT BEEN CALLED TO OUR ATTENTION ON THIS PROOF.
Material Type:	Clear	Bag Size:	5" x 7.25" + .25" Skirt	Artist:	DK	 Automated PACKAGING SYSTEMS 13555 McCracken Rd., Garfield Hts., Ohio 44125		
Notes:								

Automated Packaging Systems Inc - CONFIDENTIAL

Time/Date:

	Project No.	Artwork Number	Description	Country
	PS000002	PAA118266	WNTHRP CPJT HOLDERS	United States
	Dimensions		Drawing No.	SKU No.
5" x 7.25" with a .25" skirt		DWG-50058596	F000204491	Bag
Additional Info:		Colors:		<input checked="" type="checkbox"/> EDITOR'S COPY GS: S.Lawless DATE: 25-Dec-2018
NA		<input type="checkbox"/> Black <input type="checkbox"/> Yellow <input type="checkbox"/> Cyan <input type="checkbox"/> Magenta <input type="checkbox"/> Blue PMS 549 <input type="checkbox"/> Dieline <input type="checkbox"/> No Varnish		
Mgr	M. Guerin	Rev	GA	PR
GS	S.Lawless	1		
GA				
		CHANGES OK	CHANGES OK	CHANGES OK